

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110520\
 Data File : VW017089.D
 Acq On : 05 Nov 2020 14:46
 Operator : SY/VA
 Sample : L4615-18
 Misc : 6.13G/10ML/MSVOA W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG486

Quant Time: Nov 06 04:28:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 06 04:09:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	688527	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	580009	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	239039	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	139240	16.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.24%
7) Chloroethane-d5	2.89	69	104809	15.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.72%
10) 1,1-Dichloroethene-d2	4.02	63	179134	11.55	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.20%
20) 2-Butanone-d5	7.08	46	41202	21.24	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	42.48%
24) Chloroform-d	7.65	84	242509	15.09	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	60.36%
26) 1,2-Dichloroethane-d4	8.31	65	112587	14.09	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	56.36%#
29) Benzene-d6	8.27	84	501429	16.98	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	67.92%
33) 1,2-Dichloropropane-d6	9.27	67	135426	16.58	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	66.32%#
37) Toluene-d8	10.33	98	466784	16.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.04%
38) trans-1,3-Dichloropropene-	10.58	79	50963	13.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	55.36%
39) 2-Hexanone-d5	10.93	63	32355	22.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	45.38%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	91016	12.79	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	51.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	131642	15.63	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	62.52%#

Target Compounds

					Ovalue
13) Acetone	4.12	43	5186	3.994 ug/L	99
16) Methylene chloride	4.92	84	8180	0.881 ug/L	97
47) Tetrachloroethene	10.87	164	6555	0.852 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110520\
Data File : VW017089.D
Acq On : 05 Nov 2020 14:46
Operator : SY/VA
Sample : L4615-18
Misc : 6.13G/10ML/MSVOA W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG486

Quant Time: Nov 06 04:28:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 06 04:09:24 2020
Response via : Initial Calibration

