

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110524\
 Data File : VW030929.D
 Acq On : 05 Nov 2024 11:26
 Operator : SY/MD
 Sample : P4585-09
 Misc : 5.10g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Nov 06 00:30:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Nov 06 00:11:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	445478	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	418072	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	197044	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	162584	22.253	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.000%
7) Chloroethane-d5	2.893	69	130016	23.202	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.800%
11) 1,1-Dichloroethene-d2	4.021	65	71897	23.848	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	95.400%
21) 2-Butanone-d5	7.075	46	68640	57.088	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.180%
24) Chloroform-d	7.648	84	340252	25.909	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	103.640%
26) 1,2-Dichloroethane-d4	8.300	65	179930	27.556	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	110.240%
32) Benzene-d6	8.270	84	669143	25.763	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.040%
36) 1,2-Dichloropropane-d6	9.270	67	201060	26.260	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.040%
41) Toluene-d8	10.318	98	607605	26.081	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.320%
43) trans-1,3-Dichloroprop...	10.574	79	80887	26.063	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.240%
47) 2-Hexanone-d5	10.922	63	56321	56.758	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.520%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	150077	28.134	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.520%
66) 1,2-Dichlorobenzene-d4	13.848	152	199568	27.804	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	111.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110524\
Data File : VW030929.D
Acq On : 05 Nov 2024 11:26
Operator : SY/MD
Sample : P4585-09
Misc : 5.10g/10mL/MSVOA_W/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Nov 06 00:30:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Nov 06 00:11:36 2024
Response via : Initial Calibration

