

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110618\
 Data File : VW006698.D
 Acq On : 06 Nov 2018 20:26
 Operator : SY/AP
 Sample : J5812-25
 Misc : 5.89G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP4-A

Quant Time: Nov 07 14:39:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 01 03:48:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	246764	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	390066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	350601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	170502	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	118694	55.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.92%	
35) Dibromofluoromethane	7.89	113	125474	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
50) Toluene-d8	10.33	98	455391	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	12.62	95	149463	44.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.48%	
Target Compounds						
8) Diethyl Ether	3.68	74	3158	2.960	ug/l	89
16) Acetone	4.15	43	10078	29.428	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110618\
Data File : VW006698.D
Acq On : 06 Nov 2018 20:26
Operator : SY/AP
Sample : J5812-25
Misc : 5.89G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP4-A

Quant Time: Nov 07 14:39:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 01 03:48:08 2018
Response via : Initial Calibration

