

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110618\
 Data File : VW006713.D
 Acq On : 07 Nov 2018 04:37
 Operator : SY/AP
 Sample : J5764-13RE
 Misc : 5.07G/5ML/MSVOA W/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-110218-GRE

Quant Time: Nov 07 12:34:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 01 03:48:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	207705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	287937	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	134365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	63398	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	85037	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	7.88	113	86485	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
50) Toluene-d8	10.32	98	96714	14.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	28.12%	
62) 4-Bromofluorobenzene	12.62	95	56031	22.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.94%	
Target Compounds						
16) Acetone	4.15	43	6751	23.420	ug/l	99
20) Methylene Chloride	4.92	84	13497	1.422	ug/l	96
52) Toluene	10.39	92	5260	1.090	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110618\
Data File : VW006713.D
Acq On : 07 Nov 2018 04:37
Operator : SY/AP
Sample : J5764-13RE
Misc : 5.07G/5ML/MSVOA W/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-03-110218-GRE

Quant Time: Nov 07 12:34:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 01 03:48:08 2018
Response via : Initial Calibration

