

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110618\
 Data File : VW006728.D
 Acq On : 07 Nov 2018 11:05
 Operator : SY/AP
 Sample : J5815-03RE
 Misc : 5.25G/5ML/MSVOA W/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HL-R47183-RTRE

Quant Time: Nov 07 13:05:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 01 03:48:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	165568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	203908	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	91602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	11953	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	60162	42.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.54%	
35) Dibromofluoromethane	7.89	113	67275	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
50) Toluene-d8	10.33	98	160612	32.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.94%	
62) 4-Bromofluorobenzene	12.62	95	20286	11.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	22.98%	
Target Compounds						
16) Acetone	4.14	43	5412	23.553	ug/l	93
20) Methylene Chloride	4.91	84	12540	2.780	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110618\
Data File : VW006728.D
Acq On : 07 Nov 2018 11:05
Operator : SY/AP
Sample : J5815-03RE
Misc : 5.25G/5ML/MSVOA W/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HL-R47183-RTRE

Quant Time: Nov 07 13:05:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 01 03:48:08 2018
Response via : Initial Calibration

