

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110623\
 Data File : VW027551.D
 Acq On : 06 Nov 2023 17:13
 Operator : SY/MD
 Sample : 05239-14
 Misc : 5.69g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AQ3

Quant Time: Nov 07 00:55:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Nov 07 00:52:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	647669	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	594998	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	275498	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	171769	14.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.520%
7) Chloroethane-d5	2.899	69	123335	15.291	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.160%
11) 1,1-Dichloroethene-d2	4.021	65	80888	15.827	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.320%
21) 2-Butanone-d5	7.075	46	99176	39.174	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.340%
24) Chloroform-d	7.654	84	317473	16.801	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.200%
26) 1,2-Dichloroethane-d4	8.307	65	176854	18.420	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.680%
32) Benzene-d6	8.276	84	683925	17.193	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.760%
36) 1,2-Dichloropropane-d6	9.270	67	213418	17.371	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	69.480%#
41) Toluene-d8	10.325	98	608122	17.086	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.360%
43) trans-1,3-Dichloroprop...	10.575	79	79068	16.789	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.160%
47) 2-Hexanone-d5	10.922	63	75928	40.593	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.180%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	153602	18.596	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	74.400%
66) 1,2-Dichlorobenzene-d4	13.848	152	174956	16.707	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	66.840%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110623\
Data File : VW027551.D
Acq On : 06 Nov 2023 17:13
Operator : SY/MD
Sample : 05239-14
Misc : 5.69g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AQ3

Quant Time: Nov 07 00:55:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110623SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Nov 07 00:52:48 2023
Response via : Initial Calibration

