

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110624\
 Data File : VW030947.D
 Acq On : 06 Nov 2024 17:42
 Operator : SY/MD
 Sample : P4724-07
 Misc : 6.08g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BH9D5

Quant Time: Nov 07 04:00:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 07 03:58:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	127684	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	116510	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	40669	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	52353	25.000	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.000%	
7) Chloroethane-d5	2.893	69	39724	24.733	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	= 98.920%	
11) 1,1-Dichloroethene-d2	4.027	65	22585	26.137	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 104.560%	
21) 2-Butanone-d5	7.081	46	19253	55.867	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 111.740%	
24) Chloroform-d	7.654	84	89559	23.793	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 95.160%	
26) 1,2-Dichloroethane-d4	8.307	65	47315	25.281	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 101.120%	
32) Benzene-d6	8.276	84	168239	23.243	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 92.960%	
36) 1,2-Dichloropropane-d6	9.270	67	52217	24.472	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 97.880%	
41) Toluene-d8	10.325	98	136301	20.994	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 83.960%	
43) trans-1,3-Dichloroprop...	10.581	79	17094	19.764	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 79.040%	
47) 2-Hexanone-d5	10.922	63	12690	45.889	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 91.780%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	42493	28.584	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 114.320%	
66) 1,2-Dichlorobenzene-d4	13.854	152	37241	25.139	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 100.560%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110624\
Data File : VW030947.D
Acq On : 06 Nov 2024 17:42
Operator : SY/MD
Sample : P4724-07
Misc : 6.08g/10mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BH9D5

Quant Time: Nov 07 04:00:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 07 03:58:00 2024
Response via : Initial Calibration

