

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110718\
 Data File : VW006742.D
 Acq On : 07 Nov 2018 17:50
 Operator : SY/AP
 Sample : J5819-08
 Misc : 5.31G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BES48_SB03_3.5-4.0

Quant Time: Nov 09 10:17:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 01 03:48:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	325728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	473790	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	429051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	212544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	105394	37.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.28%	
35) Dibromofluoromethane	7.88	113	108805	36.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.18%	
50) Toluene-d8	10.33	98	387656	34.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.50%	
62) 4-Bromofluorobenzene	12.62	95	130342	31.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	63.52%	
Target Compounds						
16) Acetone	4.14	43	19426	42.973	ug/l	96
17) Carbon Disulfide	4.37	76	10900	1.442	ug/l #	94
20) Methylene Chloride	4.91	84	23429	2.299	ug/l	97
40) Benzene	8.32	78	18192m	1.537	ug/l	
52) Toluene	10.39	92	82783	10.429	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110718\
Data File : VW006742.D
Acq On : 07 Nov 2018 17:50
Operator : SY/AP
Sample : J5819-08
Misc : 5.31G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BES48_SB03_3.5-4.0

Quant Time: Nov 09 10:17:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W110218S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 01 03:48:08 2018
Response via : Initial Calibration

