

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
 Data File : VW020815.D
 Acq On : 08 Nov 2021 16:56
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 09 07:13:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 08 04:39:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	156879	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	144973	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	78179	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	37391	25.350	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.400%
7) Chloroethane-d5	2.818	69	116	0.138	ug/L	-0.07
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.560%#
11) 1,1-Dichloroethene-d2	4.013	63	49410	15.959	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.840%
21) 2-Butanone-d5	7.074	46	21153	43.182	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.360%
24) Chloroform-d	7.653	84	76308	21.891	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.560%
26) 1,2-Dichloroethane-d4	8.305	65	42181	22.731	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.920%
32) Benzene-d6	8.274	84	153516	23.438	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.760%
36) 1,2-Dichloropropane-d6	9.274	67	45328	22.352	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.400%
41) Toluene-d8	10.323	98	147927	23.214	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.840%
43) trans-1,3-Dichloroprop...	10.579	79	21635	21.509	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.040%
47) 2-Hexanone-d5	10.920	63	17242	42.021	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	168.080%#
56) 1,1,2,2-Tetrachloroeth...	12.694	84	35137	20.907	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.640%
66) 1,2-Dichlorobenzene-d4	13.853	152	56714	23.084	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.320%
Target Compounds						
13) Acetone	4.135	43	9030	19.808	ug/L	92
20) cis-1,2-Dichloroethene	7.171	96	4506	2.004	ug/L	82
39) cis-1,3-Dichloropropene	10.036	75	1840	0.550	ug/L #	65
42) Toluene	10.390	91	15051	1.612	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
Data File : VW020815.D
Acq On : 08 Nov 2021 16:56
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 09 07:13:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 08 04:39:29 2021
Response via : Initial Calibration









#13
Acetone
Concen: 19.808 ug/L
RT: 4.135 min Scan# 366
Delta R.T. 0.012 min
Lab File: VW020815.D
Acq: 08 Nov 2021 16:56




Instrument :
MSVOA_W
ClientSampleId :


Tgt Ion: 43 Resp: 9030
Ion Ratio Lower Upper
43 100
58 30.6 0.0 52.8



#20
cis-1,2-Dichloroethene
Concen: 2.004 ug/L
RT: 7.171 min Scan# 864
Delta R.T. -0.000 min
Lab File: VW020815.D
Acq: 08 Nov 2021 16:56

Tgt Ion: 96 Resp: 4506
Ion Ratio Lower Upper
96 100
61 125.6 103.7 192.7
68 0.0 0.0 0.0





#39

cis-1,3-Dichloropropene

Concen: 0.550 ug/L

RT: 10.036 min Scan# 11

Delta R.T. -0.037 min

Lab File: VW020815.D

Acq: 08 Nov 2021 16:56

Instrument :

MSVOA_W





ClientSampleId :

Tgt Ion: 75 Resp: 1840

Ion Ratio Lower Upper

75 100

77 50.7 21.8 40.6#



#42

Toluene

Concen: 1.612 ug/L

RT: 10.390 min Scan# 1392

Delta R.T. -0.000 min

Lab File: VW020815.D

Acq: 08 Nov 2021 16:56

Tgt Ion: 91 Resp: 15051

Ion Ratio Lower Upper

91 100

92 60.5 41.7 77.5

