

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
 Data File : VW020802.D
 Acq On : 08 Nov 2021 11:30
 Operator : SY/VA
 Sample : M4511-09
 Misc : 5.05g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Nov 15 00:39:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 15 00:29:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	126083	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	113158	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	60713	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	35265	29.748	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	119.000%
7) Chloroethane-d5	2.885	69	19750	29.259	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	117.040%
11) 1,1-Dichloroethene-d2	4.013	63	44843	18.022	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.080%
21) 2-Butanone-d5	7.074	46	19598	49.780	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.560%
24) Chloroform-d	7.647	84	66156	23.614	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.440%
26) 1,2-Dichloroethane-d4	8.305	65	36524	24.490	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.960%
32) Benzene-d6	8.275	84	133650	26.141	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.560%
36) 1,2-Dichloropropane-d6	9.274	67	38817	24.523	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.080%
41) Toluene-d8	10.323	98	130254	26.187	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.760%
43) trans-1,3-Dichloroprop...	10.573	79	18580	23.666	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.680%
47) 2-Hexanone-d5	10.920	63	16153	50.435	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.860%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	32048	24.430	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.720%
66) 1,2-Dichlorobenzene-d4	13.847	152	49405	25.894	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
Data File : VW020802.D
Acq On : 08 Nov 2021 11:30
Operator : SY/VA
Sample : M4511-09
Misc : 5.05g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Nov 15 00:39:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 15 00:29:21 2021
Response via : Initial Calibration

