

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
 Data File : VW020811.D
 Acq On : 08 Nov 2021 15:20
 Operator : SY/VA
 Sample : VIBLK482
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK482

Quant Time: Nov 15 01:40:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 15 00:29:21 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	151148	25.000	ug/L	# 0.00
28) Chlorobenzene-d5		11.628	117	136221	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.554	152	72879	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.349	65	40660	28.611	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	114.440%	
7) Chloroethane-d5		2.879	69	22393	27.673	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.680%	
11) 1,1-Dichloroethene-d2		4.013	63	51615	17.304	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.200%	
21) 2-Butanone-d5		7.074	46	23345	49.464	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.920%	
24) Chloroform-d		7.653	84	81937	24.397	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.600%	
26) 1,2-Dichloroethane-d4		8.305	65	45151	25.254	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.000%	
32) Benzene-d6		8.275	84	165214	26.844	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.360%	
36) 1,2-Dichloropropane-d6		9.274	67	48303	25.349	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.400%	
41) Toluene-d8		10.323	98	161300	26.939	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.760%	
43) trans-1,3-Dichloroprop...		10.579	79	23559	24.927	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.720%	
47) 2-Hexanone-d5		10.921	63	19598	50.831	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.660%	
56) 1,1,2,2-Tetrachloroeth...		12.688	84	39558	25.049	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.200%	
66) 1,2-Dichlorobenzene-d4		13.853	152	61008	26.637	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.560%	
Target Compounds							
42) Toluene		10.384	91	9895	1.128	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
Data File : VW020811.D
Acq On : 08 Nov 2021 15:20
Operator : SY/VA
Sample : VIBLK482
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK482

Quant Time: Nov 15 01:40:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 15 00:29:21 2021
Response via : Initial Calibration

