

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
 Data File : VW020816.D
 Acq On : 08 Nov 2021 17:20
 Operator : SY/VA
 Sample : VIBLK484
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK485

Quant Time: Nov 09 07:13:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 08 04:39:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	134690	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	121834	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	68146	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	37464	29.584	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 118.320%			
7) Chloroethane-d5	2.879	69	20832	28.890	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 115.560%			
11) 1,1-Dichloroethene-d2	4.019	63	48226	18.143	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 72.560%			
21) 2-Butanone-d5	7.074	46	19162	45.562	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 91.120%			
24) Chloroform-d	7.653	84	75519	25.234	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 100.920%			
26) 1,2-Dichloroethane-d4	8.311	65	39698	24.917	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 99.680%			
32) Benzene-d6	8.275	84	147877	26.864	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 107.440%			
36) 1,2-Dichloropropane-d6	9.274	67	43636	25.604	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 102.400%			
41) Toluene-d8	10.323	98	144877	27.053	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 108.200%			
43) trans-1,3-Dichloroprop...	10.573	79	20645	24.423	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 97.680%			
47) 2-Hexanone-d5	10.920	63	15533	45.045	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 180.200%#			
56) 1,1,2,2-Tetrachloroeth...	12.694	84	33309	23.583	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 94.320%			
66) 1,2-Dichlorobenzene-d4	13.853	152	53522	24.992	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 99.960%			
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	7.171	96	1367	0.708	ug/L	95
42) Toluene	10.390	91	8073	1.029	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110821\
Data File : VW020816.D
Acq On : 08 Nov 2021 17:20
Operator : SY/VA
Sample : VIBLK484
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK485

Quant Time: Nov 09 07:13:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 08 04:39:29 2021
Response via : Initial Calibration

