

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110921\
 Data File : VW020822.D
 Acq On : 09 Nov 2021 09:41
 Operator : SY/VA
 Sample : VW1109SBL01
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK404

Quant Time: Nov 15 03:01:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 15 03:01:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	142005	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	127355	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	73172	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	36238	27.141	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.560%	
7) Chloroethane-d5	2.885	69	20013	26.324	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 105.280%	
11) 1,1-Dichloroethene-d2	4.013	63	44563	15.901	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 63.600%	
21) 2-Butanone-d5	7.074	46	20597	46.451	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 92.900%	
24) Chloroform-d	7.647	84	73073	23.159	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 92.640%	
26) 1,2-Dichloroethane-d4	8.305	65	40761	24.266	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 97.080%	
32) Benzene-d6	8.275	84	145490	25.285	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 101.160%	
36) 1,2-Dichloropropane-d6	9.274	67	42332	23.762	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 95.040%	
41) Toluene-d8	10.323	98	142217	25.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 101.640%	
43) trans-1,3-Dichloroprop...	10.579	79	21030	23.800	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 95.200%	
47) 2-Hexanone-d5	10.921	63	17569	48.741	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 97.480%	
56) 1,1,2,2-Tetrachloroeth...	12.695	84	35682	24.168	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 96.680%	
66) 1,2-Dichlorobenzene-d4	13.853	152	56602	24.614	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 98.440%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110921\
Data File : VW020822.D
Acq On : 09 Nov 2021 09:41
Operator : SY/VA
Sample : VW1109SBL01
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK404

Quant Time: Nov 15 03:01:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 15 03:01:39 2021
Response via : Initial Calibration

