

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110921\
 Data File : VW020823.D
 Acq On : 09 Nov 2021 15:34
 Operator : SY/VA
 Sample : M4464-21
 Misc : 5.02g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Nov 15 03:03:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Nov 15 03:01:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	142392	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	128451	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	69111	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	38768	28.957	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 115.840%	
7) Chloroethane-d5	2.891	69	25065	32.880	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 131.520%	
11) 1,1-Dichloroethene-d2	4.019	63	51725	18.407	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 73.640%	
21) 2-Butanone-d5	7.074	46	19287	43.379	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 86.760%	
24) Chloroform-d	7.653	84	79124	25.008	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 100.040%	
26) 1,2-Dichloroethane-d4	8.305	65	42331	25.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 100.520%	
32) Benzene-d6	8.275	84	158142	27.249	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 109.000%	
36) 1,2-Dichloropropane-d6	9.274	67	45674	25.419	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 101.680%	
41) Toluene-d8	10.323	98	154364	27.340	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 109.360%	
43) trans-1,3-Dichloroprop...	10.579	79	21509	24.135	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 96.520%	
47) 2-Hexanone-d5	10.920	63	16104	44.295	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 88.600%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	33936	22.789	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 91.160%	
66) 1,2-Dichlorobenzene-d4	13.853	152	58530	26.949	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 107.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110921\
Data File : VW020823.D
Acq On : 09 Nov 2021 15:34
Operator : SY/VA
Sample : M4464-21
Misc : 5.02g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Nov 15 03:03:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110621SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Nov 15 03:01:39 2021
Response via : Initial Calibration

