

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW111119\
Quantitation Report (Qedit)




Data File : VW013967.D
Acq On : 11 Nov 2019 18:52
Operator : SY/VA
Sample : K5784-04
Misc : 4.89G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampled :
JLNB1

Manual Integrations
APPROVED

MMDadoda
11/12/2019 4:18:54 PM

Quant Time: Nov 12 06:14:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 06:10:10 2019
Response via : Initial Calibration



TIC: VW013967.D

(13) Acetone (T)

4.129min (+0.012) 2.33ug/L

response 3347

Ion	Exp%	Act%
43.00	100	100
58.00	32.70	11.98
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW111119\
Quantitation Report (Qedit)




Data File : VW013967.D
Acq On : 11 Nov 2019 18:52
Operator : SY/VA
Sample : K5784-04
Misc : 4.89G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JLNB1

Manual Integrations
APPROVED

MMDadoda
11/12/2019 4:18:54 PM

Quant Time: Nov 12 06:14:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 06:10:10 2019
Response via : Initial Calibration



(13) Acetone (T)

4.129min (+0.012) 2.57ug/L m

response 3686

Ion	Exp%	Act%
43.00	100	100
58.00	32.70	10.88
0.00	0.00	0.00
0.00	0.00	0.00

11/12/1954

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW111119\
Quantitation Report (QT Reviewed)

Data File : VW013967.D
Acq On : 11 Nov 2019 18:52
Operator : SY/VA
Sample : K5784-04
Misc : 4.89G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JLNB1

Manual Integrations
APPROVED

MMDadoda
11/12/2019 4:18:54 PM

Quant Time: Nov 12 07:06:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 06:10:10 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	303941	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	279272	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	125422	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	104538	21.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.24%
7) Chloroethane-d5	2.89	69	94538	24.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.20%
10) 1,1-Dichloroethene-d2	4.01	63	175001	17.58	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.32%
20) 2-Butanone-d5	7.07	46	73317	52.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.08%
24) Chloroform-d	7.65	84	248507	26.59	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.36%
26) 1,2-Dichloroethane-d4	8.30	65	133489	28.41	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	113.64%
29) Benzene-d6	8.27	84	538675	27.19	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.76%
33) 1,2-Dichloropropane-d6	9.27	67	165983	28.60	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	114.40%
37) Toluene-d8	10.32	98	487361	25.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	103.76%
38) trans-1,3-Dichloropropene-	10.57	79	59694	24.01	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.04%
39) 2-Hexanone-d5	10.92	63	58873	53.49	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.98%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	122571	27.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.84%
61) 1,2-Dichlorobenzene-d4	13.85	152	138670	27.52	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	110.08%

Target Compounds

13) Acetone	4.13	43	3686m)	2.571	ug/L	Qvalue
-------------	------	----	--------	-------	------	--------

11/12/19 SY

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW111119\
Quantitation Report (QT Reviewed)

Data File : VW013967.D
Acq On : 11 Nov 2019 18:52
Operator : SY/VA
Sample : K5784-04
Misc : 4.89G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JLNB1

Manual Integrations
APPROVED

MMDadoda
11/12/2019 4:18:54 PM

Quant Time: Nov 12 07:06:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 12 06:10:10 2019
Response via : Initial Calibration

