

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW111218\
 Data File : VW006829.D
 Acq On : 13 Nov 2018 02:44
 Operator : SY/AP
 Sample : J5909-03
 Misc : 6.61G/5ML/MSVOA_W/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-15-S

Quant Time: Nov 13 04:06:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 14:58:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	330894	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	438016	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	386992	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	156473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	119493	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
35) Dibromofluoromethane	7.88	113	114430	43.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.74%	
50) Toluene-d8	10.33	98	443840	41.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.12%	
62) 4-Bromofluorobenzene	12.62	95	132869	33.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.74%	
Target Compounds						
8) Diethyl Ether	3.69	74	3592	2.805	ug/l	88
16) Acetone	4.14	43	40906	117.752	ug/l	98
20) Methylene Chloride	4.92	84	26380	3.507	ug/l	94
25) 2-Butanone	7.18	43	14220	24.204	ug/l	93
95) Naphthalene	15.37	128	11225	2.045	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW111218\
Data File : VW006829.D
Acq On : 13 Nov 2018 02:44
Operator : SY/AP
Sample : J5909-03
Misc : 6.61G/5ML/MSVOA_W/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-15-S

Quant Time: Nov 13 04:06:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 14:58:58 2018
Response via : Initial Calibration

