

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111318\
 Data File : VW006851.D
 Acq On : 13 Nov 2018 13:55
 Operator : SY/AP
 Sample : J5887-11RE
 Misc : 4.39G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 W-2RE

Quant Time: Nov 13 15:48:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 14:58:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	111730	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	152750	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	130505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	51741	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	37073	45.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.78%	
35) Dibromofluoromethane	7.89	113	39313	43.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.44%	
50) Toluene-d8	10.33	98	145156	38.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.00%	
62) 4-Bromofluorobenzene	12.62	95	43167	31.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	62.18%	
Target Compounds						
16) Acetone	4.13	43	3814	32.515	ug/l	94
20) Methylene Chloride	4.93	84	8127	2.692	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111318\
Data File : VW006851.D
Acq On : 13 Nov 2018 13:55
Operator : SY/AP
Sample : J5887-11RE
Misc : 4.39G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
W-2RE

Quant Time: Nov 13 15:48:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 14:58:58 2018
Response via : Initial Calibration

