

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111318\
 Data File : VW006865.D
 Acq On : 13 Nov 2018 19:52
 Operator : SY/AP
 Sample : J5895-01
 Misc : 10.23G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1

Manual Integrations
 APPROVED

apatel
 11/14/2018 4:37:10 PM

Quant Time: Nov 14 06:50:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 14:58:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	225369	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	300338	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	277175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	139841	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	64532	39.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.34%	
35) Dibromofluoromethane	7.89	113	69471	38.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.68%	
50) Toluene-d8	10.33	98	292451	39.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.90%	
62) 4-Bromofluorobenzene	12.62	95	103371	37.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.72%	
Target Compounds						
					Qvalue	
16) Acetone	4.14	43	7231	30.561	ug/l #	87
20) Methylene Chloride	4.92	84	14039	1.472	ug/l	89
39) Methylcyclohexane	9.34	83	8857	2.429	ug/l	90
64) Tetrachloroethene	10.87	164	3234	1.456	ug/l	87
67) Ethyl Benzene	11.73	91	13091	1.249	ug/l	100
68) m/p-Xylenes	11.85	106	6115	1.426	ug/l	93
78) n-propylbenzene	12.81	91	30464	2.659	ug/l	100
80) 1,3,5-Trimethylbenzene	12.95	105	18781	2.251	ug/l	99
84) 1,2,4-Trimethylbenzene	13.26	105	67602	7.961	ug/l	96
89) n-Butylbenzene	13.83	91	11531m	1.327	ug/l	
95) Naphthalene	15.38	128	21288	4.340	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111318\
Data File : VW006865.D
Acq On : 13 Nov 2018 19:52
Operator : SY/AP
Sample : J5895-01
Misc : 10.23G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1

Manual Integrations
APPROVED

apatel
11/14/2018 4:37:10 PM

Quant Time: Nov 14 06:50:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W111218S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 14:58:58 2018
Response via : Initial Calibration

