


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW111319\
 Data File : VW013994.D
 Acq On : 13 Nov 2019 08:44
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5.0ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W102319S.M
 Title : SW846 8260
 Last Update : Wed Oct 23 14:11:46 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	34346	PASS
75	95	30	60	47.6	96277	PASS
95	95	100	100	100.0	202346	PASS
96	95	5	9	6.7	13654	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.9	177813	PASS
175	174	5	9	5.2	9196	PASS
176	174	95	101	95.9	170472	PASS
177	176	5	9	7.2	12323	PASS