

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111320\
 Data File : VW017207.D
 Acq On : 13 Nov 2020 13:05
 Operator : SY/VA
 Sample : VIBLK07
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK07

Quant Time: Nov 17 04:29:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 16 11:48:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	424117	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	372035	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	185522	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	114144	21.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.48%
7) Chloroethane-d5	2.89	69	96661	23.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.92%
10) 1,1-Dichloroethene-d2	4.02	63	163291	17.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.36%
20) 2-Butanone-d5	7.08	46	50855	42.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.14%
24) Chloroform-d	7.65	84	247929	25.04	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.16%
26) 1,2-Dichloroethane-d4	8.31	65	126483	25.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.80%
29) Benzene-d6	8.27	84	473401	24.99	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.96%
33) 1,2-Dichloropropane-d6	9.27	67	131613	25.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.48%
37) Toluene-d8	10.33	98	447933	25.07	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.28%
38) trans-1,3-Dichloropropene-	10.58	79	54691	23.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.60%
39) 2-Hexanone-d5	10.93	63	40851	44.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.34%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	108479	23.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.04%
61) 1,2-Dichlorobenzene-d4	13.85	152	172776	26.43	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.72%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	18719	3.274 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111320\
Data File : VW017207.D
Acq On : 13 Nov 2020 13:05
Operator : SY/VA
Sample : VIBLK07
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK07

Quant Time: Nov 17 04:29:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 16 11:48:24 2020
Response via : Initial Calibration

