

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111320\
 Data File : VW017209.D
 Acq On : 13 Nov 2020 13:52
 Operator : SY/VA
 Sample : L4693-11
 Misc : 5.01G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK01

Quant Time: Nov 17 04:46:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 16 11:48:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	397933	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	367850	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	177700	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	102115	20.37	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.48%
7) Chloroethane-d5	2.89	69	89345	23.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.52%
10) 1,1-Dichloroethene-d2	4.02	63	140783	15.70	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.80%
20) 2-Butanone-d5	7.09	46	53050	47.33	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.66%
24) Chloroform-d	7.65	84	217142	23.37	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.48%
26) 1,2-Dichloroethane-d4	8.31	65	115446	25.00	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.00%
29) Benzene-d6	8.27	84	411832	21.98	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.92%
33) 1,2-Dichloropropane-d6	9.27	67	114893	22.17	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.68%
37) Toluene-d8	10.33	98	387667	21.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.80%
38) trans-1,3-Dichloropropene-	10.58	79	50092	21.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.76%
39) 2-Hexanone-d5	10.93	63	44260	48.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.90%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	109062	24.16	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.64%
61) 1,2-Dichlorobenzene-d4	13.85	152	155548	24.85	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.40%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	20746	3.867 ug/L	96
43) 4-Methyl-2-pentanone	10.22	43	1959	0.703 ug/L #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111320\
Data File : VW017209.D
Acq On : 13 Nov 2020 13:52
Operator : SY/VA
Sample : L4693-11
Misc : 5.01G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK01

Quant Time: Nov 17 04:46:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 16 11:48:24 2020
Response via : Initial Calibration

