

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111324\
 Data File : VW031006.D
 Acq On : 13 Nov 2024 11:55
 Operator : SY/MD
 Sample : VW1113SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK595

Quant Time: Nov 14 00:31:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Nov 14 00:30:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	408203	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	387333	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	169072	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	106119	15.851	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.400%
7) Chloroethane-d5	2.899	69	90390	17.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.400%
11) 1,1-Dichloroethene-d2	4.021	65	53932	19.523	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.080%
21) 2-Butanone-d5	7.075	46	52019	47.215	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.440%
24) Chloroform-d	7.648	84	272895	22.677	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.720%
26) 1,2-Dichloroethane-d4	8.307	65	134389	22.461	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.840%
32) Benzene-d6	8.276	84	526970	21.900	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	9.276	67	158223	22.305	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.200%
41) Toluene-d8	10.325	98	475043	22.009	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.040%
43) trans-1,3-Dichloroprop...	10.575	79	57870	20.126	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.520%
47) 2-Hexanone-d5	10.922	63	43650	47.480	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.960%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	120854	24.454	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	13.849	152	160893	26.125	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111324\
Data File : VW031006.D
Acq On : 13 Nov 2024 11:55
Operator : SY/MD
Sample : VW1113SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK595

Quant Time: Nov 14 00:31:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM110424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Nov 14 00:30:57 2024
Response via : Initial Calibration

