

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111521\
 Data File : VW020896.D
 Acq On : 15 Nov 2021 17:08
 Operator : SY/VA
 Sample : VIBLK492
 Misc : 5.00g/10.00mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK492

Quant Time: Nov 16 04:20:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Nov 16 04:07:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	127288	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	119680	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	64545	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	33067	22.396	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.600%
7) Chloroethane-d5	2.885	69	19108	20.066	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.280%
11) 1,1-Dichloroethene-d2	4.019	63	43261	15.144	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.560%
21) 2-Butanone-d5	7.074	46	16000	43.416	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.840%
24) Chloroform-d	7.647	84	68114	21.238	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.960%
26) 1,2-Dichloroethane-d4	8.305	65	35649	21.898	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	8.275	84	134990	21.401	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	9.274	67	38292	21.319	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.280%
41) Toluene-d8	10.323	98	133597	21.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.360%
43) trans-1,3-Dichloroprop...	10.579	79	17886	21.067	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.280%
47) 2-Hexanone-d5	10.920	63	13150	44.142	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.280%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	29977	21.786	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.160%
66) 1,2-Dichlorobenzene-d4	13.847	152	51723	22.484	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111521\
Data File : VW020896.D
Acq On : 15 Nov 2021 17:08
Operator : SY/VA
Sample : VIBLK492
Misc : 5.00g/10.00mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK492

Quant Time: Nov 16 04:20:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Nov 16 04:07:18 2021
Response via : Initial Calibration

