

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112018\
 Data File : VW006952.D
 Acq On : 19 Nov 2018 19:43
 Operator : SY/AP
 Sample : J5952-05
 Misc : 3.91G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VTW-03(4-6)

Quant Time: Nov 20 05:41:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W111718S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 17 04:45:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	92490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	142159	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	110065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	32182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	44750	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
35) Dibromofluoromethane	7.89	113	40899	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	10.33	98	146277	40.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.78%	
62) 4-Bromofluorobenzene	12.62	95	40731	29.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.40%	
Target Compounds						
16) Acetone	4.14	43	10745	68.152	ug/l	94
17) Carbon Disulfide	4.39	76	12177	4.517	ug/l	98
20) Methylene Chloride	4.92	84	6281	2.584	ug/l	97
95) Naphthalene	15.37	128	1817	1.517	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112018\
Data File : VW006952.D
Acq On : 19 Nov 2018 19:43
Operator : SY/AP
Sample : J5952-05
Misc : 3.91G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VTW-03(4-6)

Quant Time: Nov 20 05:41:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W111718S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 17 04:45:18 2018
Response via : Initial Calibration

