


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112118\
Data File : VW006972.D
Acq On : 20 Nov 2018 12:49
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Title : SW846 8260
Last Update : Wed Nov 21 03:23:02 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	14158	PASS
75	95	30	60	48.3	39016	PASS
95	95	100	100	100.0	80821	PASS
96	95	5	9	6.6	5366	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.1	57456	PASS
175	174	5	9	7.9	4519	PASS
176	174	95	101	95.9	55080	PASS
177	176	5	9	7.1	3908	PASS