

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112124\
 Data File : VW031081.D
 Acq On : 21 Nov 2024 14:56
 Operator : SY/MD
 Sample : P4942-01
 Misc : 8.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F7H70

Quant Time: Nov 22 00:08:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM112024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Nov 22 00:06:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	575349	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	506808	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	193504	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	90807	12.947	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	51.800%
7) Chloroethane-d5	2.899	69	81984	14.810	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.240%
11) 1,1-Dichloroethene-d2	4.021	65	44246	12.767	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.080%
21) 2-Butanone-d5	7.075	46	76811	47.563	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.120%
24) Chloroform-d	7.648	84	256642	16.421	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	65.680%
26) 1,2-Dichloroethane-d4	8.307	65	139092	18.215	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	72.840%
32) Benzene-d6	8.276	84	458581	15.978	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	63.920%
36) 1,2-Dichloropropane-d6	9.270	67	151666	17.921	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.680%
41) Toluene-d8	10.325	98	375889	14.376	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	57.520%
43) trans-1,3-Dichloroprop...	10.575	79	58482	16.599	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.400%
47) 2-Hexanone-d5	10.922	63	58414	48.026	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.060%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	142470	22.103	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.400%
66) 1,2-Dichlorobenzene-d4	13.849	152	117940	17.439	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.760%#
Target Compounds					Qvalue	
46) Tetrachloroethene	10.861	164	3458	0.492	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112124\
Data File : VW031081.D
Acq On : 21 Nov 2024 14:56
Operator : SY/MD
Sample : P4942-01
Misc : 8.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F7H70

Quant Time: Nov 22 00:08:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM112024SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Nov 22 00:06:47 2024
Response via : Initial Calibration

