

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112218\
 Data File : VW007015.D
 Acq On : 21 Nov 2018 21:47
 Operator : VAY/AP
 Sample : J5772-03RE
 Misc : 5.43G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PL-01-112018-ARE

Quant Time: Nov 22 05:53:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	7854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	11888	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	5550	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	2937	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.32	65	4327	56.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.54%	
35) Dibromofluoromethane	7.88	113	3855	57.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.84%	
50) Toluene-d8	10.33	98	2775	9.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	18.72%	
62) 4-Bromofluorobenzene	12.62	95	3547	30.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.80%	
Target Compounds						
16) Acetone	4.13	43	1236	110.714	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112218\
Data File : VW007015.D
Acq On : 21 Nov 2018 21:47
Operator : VAY/AP
Sample : J5772-03RE
Misc : 5.43G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PL-01-112018-ARE

Quant Time: Nov 22 05:53:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

