

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112218\
 Data File : VW007019.D
 Acq On : 22 Nov 2018 01:36
 Operator : VAY/AP
 Sample : VW1122SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1122SBL02

Quant Time: Nov 22 05:59:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	106374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	169884	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	163251	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	78381	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	59337	56.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.96%	
35) Dibromofluoromethane	7.88	113	51344	53.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.04%	
50) Toluene-d8	10.32	98	226473	53.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.92%	
62) 4-Bromofluorobenzene	12.62	95	87954	52.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112218\
Data File : VW007019.D
Acq On : 22 Nov 2018 01:36
Operator : VAY/AP
Sample : VW1122SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1122SBL02

Quant Time: Nov 22 05:59:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

