


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW112218\
Data File : VW006993.D
Acq On : 21 Nov 2018 10:52
Operator : VAY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Title : SW846 8260
Last Update : Wed Nov 21 03:23:02 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	17154	PASS
75	95	30	60	49.4	48765	PASS
95	95	100	100	100.0	98808	PASS
96	95	5	9	7.0	6949	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	71.0	70162	PASS
175	174	5	9	8.2	5773	PASS
176	174	95	101	96.6	67770	PASS
177	176	5	9	7.3	4949	PASS