


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW112218\
Data File : VW007017.D
Acq On : 21 Nov 2018 23:55
Operator : VAY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Title : SW846 8260
Last Update : Wed Nov 21 03:23:02 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	14374	PASS
75	95	30	60	49.2	42184	PASS
95	95	100	100	100.0	85800	PASS
96	95	5	9	6.5	5546	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.4	63010	PASS
175	174	5	9	8.3	5260	PASS
176	174	95	101	97.5	61426	PASS
177	176	5	9	6.5	3998	PASS