

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112525\
 Data File : VW032534.D
 Acq On : 25 Nov 2025 13:48
 Operator : SY/MD
 Sample : Q3719-02
 Misc : 5.28g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1125-10B

Quant Time: Nov 26 01:40:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 11 03:48:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	227021	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	528970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	565806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	258185	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	162152	55.402	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	110.800%
35) Dibromofluoromethane	7.898	113	156487	49.411	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	98.820%
50) Toluene-d8	10.325	98	539022	44.364	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	88.720%
62) 4-Bromofluorobenzene	12.617	95	235894	51.396	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	102.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112525\
Data File : VW032534.D
Acq On : 25 Nov 2025 13:48
Operator : SY/MD
Sample : Q3719-02
Misc : 5.28g/5mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1125-10B

Quant Time: Nov 26 01:40:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111025S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 11 03:48:21 2025
Response via : Initial Calibration

