

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW112618\
 Data File : VW007076.D
 Acq On : 27 Nov 2018 05:36
 Operator : VAY/AP
 Sample : J5996-25
 Misc : 5.31G/5ML/MSVOA_W/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 AOC1-06-A

Quant Time: Nov 27 07:57:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	97913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	155367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	118703	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	38557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	52936	55.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.44%	
35) Dibromofluoromethane	7.95	113	214	0.24	ug/l	0.05
Spiked Amount 50.000			Recovery	=	0.48%	
50) Toluene-d8	10.33	98	166428	42.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.92%	
62) 4-Bromofluorobenzene	12.62	95	49665	32.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.14%	
Target Compounds						
16) Acetone	4.14	43	8336	59.90	ug/l	Qvalue 100
17) Carbon Disulfide	4.38	76	6125	2.14	ug/l #	94
20) Methylene Chloride	4.92	84	64363	59.00	ug/l	98
52) Toluene	10.39	92	4184	1.50	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW112618\
Data File : VW007076.D
Acq On : 27 Nov 2018 05:36
Operator : VAY/AP
Sample : J5996-25
Misc : 5.31G/5ML/MSVOA_W/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
AOC1-06-A

Quant Time: Nov 27 07:57:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

