

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW112618\
 Data File : VW007080.D
 Acq On : 27 Nov 2018 10:38
 Operator : VAY/AP
 Sample : J5774-05
 Misc : 5.72/5ML/MSVOA_W/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SU-03-112118-A

Quant Time: Nov 27 12:39:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	124652	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	200873	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	158692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	52075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	64224	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	7.88	113	54482	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
50) Toluene-d8	10.32	98	201899	40.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.60%	
62) 4-Bromofluorobenzene	12.62	95	66743	33.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.72%	
Target Compounds						
16) Acetone	4.14	43	7287	41.13	ug/l	92
20) Methylene Chloride	4.92	84	10033	1.91	ug/l	94
52) Toluene	10.39	92	12639	3.50	ug/l	97
84) 1,2,4-Trimethylbenzene	13.25	105	7476	2.14	ug/l	97
95) Naphthalene	15.38	128	179435	96.40	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW112618\
Data File : VW007080.D
Acq On : 27 Nov 2018 10:38
Operator : VAY/AP
Sample : J5774-05
Misc : 5.72/5ML/MSVOA_W/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SU-03-112118-A

Quant Time: Nov 27 12:39:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

