

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112619\
 Data File : VW014128.D
 Acq On : 26 Nov 2019 12:04
 Operator : SY/VA
 Sample : K6007-07RE
 Misc : 4.34G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BL5RE

Quant Time: Nov 27 07:37:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 27 06:15:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	221510	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	115514	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	20000	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	58683	16.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.44%
7) Chloroethane-d5	2.89	69	61792	21.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.28%
10) 1,1-Dichloroethene-d2	4.01	63	94013	12.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.84%
20) 2-Butanone-d5	7.07	46	64575	62.89	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.78%
24) Chloroform-d	7.65	84	167878	24.65	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	8.30	65	81585	23.83	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.32%
29) Benzene-d6	8.27	84	332272	40.54	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	162.16%#
33) 1,2-Dichloropropane-d6	9.27	67	107301	44.69	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	178.76%#
37) Toluene-d8	10.32	98	218239	28.08	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	112.32%
38) trans-1,3-Dichloropropene-	10.57	79	17151	16.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.72%
39) 2-Hexanone-d5	10.92	63	35369	77.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	155.38%#
48) 1,1,1,2,2-Tetrachloroethane-	12.69	84	71264	39.31	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	157.24%#
61) 1,2-Dichlorobenzene-d4	13.85	152	23571	29.33	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	117.32%

Target Compounds

					Ovalue
13) Acetone	4.13	43	4346	4.160 ug/L	50
31) 1,1,1-Trichloroethane	7.88	97	1785	0.617 ug/L #	86
36) Methylcyclohexane	9.34	83	4058	1.001 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112619\
Data File : VW014128.D
Acq On : 26 Nov 2019 12:04
Operator : SY/VA
Sample : K6007-07RE
Misc : 4.34G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BL5RE

Quant Time: Nov 27 07:37:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 27 06:15:20 2019
Response via : Initial Calibration

