

Quantitation Report (QT/LSC Reviewed)


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
 Data File : VW020960.D
 Acq On : 26 Nov 2021 16:15
 Operator : SY/VA
 Sample : M4833-20
 Misc : 6.29g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 Client Sampled :
 ESQN7

Quant Time: Nov 27 00:58:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 27 00:54:15 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/28/2021
 Supervised By : Mahesh Dadoda 11/29/2021



Quantitation Report (Qedit)


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
 Data File : VW020960.D
 Acq On : 26 Nov 2021 16:15
 Operator : SY/VA
 Sample : M4833-20
 Misc : 6.29g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ESQN7

Quant Time: Nov 27 00:58:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 27 00:54:15 2021
 Response via : Initial Calibration

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 11/28/2021
 Supervised By :Mahesh Dadoda 11/29/2021



(13) Acetone (T)

4.123min (-0.006) 3.09 ug/L

response 926

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	26.40	44.82
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
 Data File : VW020960.D
 Acq On : 26 Nov 2021 16:15
 Operator : SY/VA
 Sample : M4833-20
 Misc : 6.29g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 ESQN7

Manual IntegrationsAPPROVED

Quant Time: Nov 27 00:58:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 27 00:54:15 2021
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 11/28/2021
 Supervised By :Mahesh Dadoda 11/29/2021



(13) Acetone (T)

4.135min (+ 0.006) 6.39 ug/L m

response 1914

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	26.40	21.68
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (QT/LSC Reviewed)

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
 Data File : VW020960.D
 Acq On : 26 Nov 2021 16:15
 Operator : SY/VA
 Sample : M4833-20
 Misc : 6.29g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 Client Sampled :
 ESQN7

Quant Time: Nov 27 00:58:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 27 00:54:15 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/28/2021
 Supervised By : Mahesh Dadoda 11/29/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	115949	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	92863	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	33894	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	25470	18.938	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	75.760%	
7) Chloroethane-d5	2.885	69	17292	19.935	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	79.760%	
11) 1,1-Dichloroethene-d2	4.013	63	30905	11.876	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	47.520%	
21) 2-Butanone-d5	7.080	46	15888	47.328	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	94.660%	
24) Chloroform-d	7.653	84	56397	19.304	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	77.200%	
26) 1,2-Dichloroethane-d4	8.305	65	32419	21.862	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	87.440%	
32) Benzene-d6	8.275	84	105446	21.544	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	86.160%	
36) 1,2-Dichloropropane-d6	9.275	67	32040	22.989	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	91.960%	
41) Toluene-d8	10.323	98	92159	18.971	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	75.880%	
43) trans-1,3-Dichloroprop...	10.579	79	13808	20.961	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	83.840%	
47) 2-Hexanone-d5	10.921	63	12955	56.046	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	112.100%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	27366	25.632	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	102.520%	
66) 1,2-Dichlorobenzene-d4	13.853	152	24794	20.524	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	82.080%	
Target Compounds						
13) Acetone	4.135	43	1914m	6.386	ug/L	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed