

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
 Data File : VW020950.D
 Acq On : 26 Nov 2021 12:14
 Operator : SY/VA
 Sample : M4833-01
 Misc : 5.32g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ESQM1

Quant Time: Nov 27 00:56:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Nov 27 00:54:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	135014	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	118390	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	59200	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	26969	17.221	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.880%
7) Chloroethane-d5	2.891	69	16555	16.390	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.560%
11) 1,1-Dichloroethene-d2	4.019	63	36389	12.009	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.040%
21) 2-Butanone-d5	7.079	46	15056	38.517	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.040%
24) Chloroform-d	7.646	84	63770	18.745	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	75.000%
26) 1,2-Dichloroethane-d4	8.305	65	34415	19.930	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	79.720%
32) Benzene-d6	8.274	84	118904	19.056	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.240%
36) 1,2-Dichloropropane-d6	9.274	67	35686	20.084	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.320%
41) Toluene-d8	10.323	98	110343	17.817	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.280%
43) trans-1,3-Dichloroprop...	10.579	79	15780	18.789	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.160%
47) 2-Hexanone-d5	10.926	63	12457	42.271	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.540%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	28394	20.860	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.440%
66) 1,2-Dichlorobenzene-d4	13.853	152	39578	18.758	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	75.040%
Target Compounds					Qvalue	
13) Acetone	4.135	43	1224	3.507	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112621\
Data File : VW020950.D
Acq On : 26 Nov 2021 12:14
Operator : SY/VA
Sample : M4833-01
Misc : 5.32g/10.0mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ESQM1

Quant Time: Nov 27 00:56:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Nov 27 00:54:15 2021
Response via : Initial Calibration





#13
 Acetone
 Concen: 3.507 ug/L
 RT: 4.135 min Scan# 30
 Delta R.T. 0.006 min
 Lab File: VW020950.D
 Acq: 26 Nov 2021 12:14

Instrument :
 MSVOA_W
 ClientSampleId :
 ESQM1

Tgt Ion: 43 Resp: 1224
 Ion Ratio Lower Upper
 43 100
 58 25.8 0.0 52.8

