

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW112718\
 Data File : VW007087.D
 Acq On : 27 Nov 2018 15:04
 Operator : VAY/AP
 Sample : J6003-03RE
 Misc : 4.90G/5ML/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 NON-UST-03R-ARE

Quant Time: Nov 28 02:40:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	52763	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	87031	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	82373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	33190	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	32767	63.43	ug/l	0.00
Spiked Amount			Recovery	=	126.86%	
35) Dibromofluoromethane	7.89	113	28753	58.50	ug/l	0.00
Spiked Amount			Recovery	=	117.00%	
50) Toluene-d8	10.33	98	107062	49.33	ug/l	0.00
Spiked Amount			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.62	95	41980	49.15	ug/l	0.00
Spiked Amount			Recovery	=	98.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.14	43	2204	29.39	ug/l #	87
20) Methylene Chloride	4.93	84	1484	Below Cal	#	75
52) Toluene	10.39	92	2211	1.42	ug/l	95
73) Isopropylbenzene	12.47	105	38400	14.89	ug/l	99
78) n-propylbenzene	12.81	91	76106	24.36	ug/l	99
83) tert-Butylbenzene	13.21	119	5465	2.83	ug/l	92
85) sec-Butylbenzene	13.39	105	97645	35.79	ug/l	87
89) n-Butylbenzene	13.83	91	53003	23.40	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW112718\
Data File : VW007087.D
Acq On : 27 Nov 2018 15:04
Operator : VAY/AP
Sample : J6003-03RE
Misc : 4.90G/5ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
NON-UST-03R-ARE

Quant Time: Nov 28 02:40:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

