

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW112718\
 Data File : VW007100.D
 Acq On : 27 Nov 2018 20:34
 Operator : VAY/AP
 Sample : J6083-05
 Misc : 5.26G/5ML/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WC-3

Quant Time: Nov 28 02:18:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	45816	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	72926	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	69952	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	31740	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	25250	56.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.58%	
35) Dibromofluoromethane	7.89	113	21744	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
50) Toluene-d8	10.33	98	83804	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
62) 4-Bromofluorobenzene	12.62	95	32769	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
Target Compounds						
52) Toluene	10.39	92	2415	1.84	ug/l	96
68) m/p-Xylenes	11.85	106	2179	2.07	ug/l	95
69) o-Xylene	12.17	106	1024	1.03	ug/l	82
84) 1,2,4-Trimethylbenzene	13.26	105	5712	2.69	ug/l	99
95) Naphthalene	15.38	128	2565	2.26	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW112718\
Data File : VW007100.D
Acq On : 27 Nov 2018 20:34
Operator : VAY/AP
Sample : J6083-05
Misc : 5.26G/5ML/MSVOA_W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WC-3

Quant Time: Nov 28 02:18:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

