


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW112718\
Data File : VW007081.D
Acq On : 27 Nov 2018 12:02
Operator : VAY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Title : SW846 8260
Last Update : Wed Nov 21 03:23:02 2018



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	15505	PASS
75	95	30	60	48.7	45256	PASS
95	95	100	100	100.0	92946	PASS
96	95	5	9	6.9	6392	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.9	68704	PASS
175	174	5	9	8.2	5651	PASS
176	174	95	101	99.4	68261	PASS
177	176	5	9	6.6	4536	PASS