

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112818\
 Data File : VW007111.D
 Acq On : 28 Nov 2018 14:17
 Operator : VAY/AP
 Sample : J6080-04
 Misc : 10.28G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 RCA-07

Quant Time: Nov 29 04:49:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	12854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	19663	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	14598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	4519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	6831	54.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.56%	
35) Dibromofluoromethane	7.89	113	5145	46.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.66%	
50) Toluene-d8	10.33	98	21223	43.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.56%	
62) 4-Bromofluorobenzene	12.62	95	6333	32.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	65.64%	
Target Compounds						
16) Acetone	4.13	43	3651	199.825	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112818\
Data File : VW007111.D
Acq On : 28 Nov 2018 14:17
Operator : VAY/AP
Sample : J6080-04
Misc : 10.28G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
RCA-07

Quant Time: Nov 29 04:49:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

