

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112818\
 Data File : VW007121.D
 Acq On : 28 Nov 2018 18:31
 Operator : VAY/AP
 Sample : J6082-01
 Misc : 7.24G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S49-0490

Quant Time: Nov 29 13:26:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	22281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	36974	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	38123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	20956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	10920	50.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.12%	
35) Dibromofluoromethane	7.89	113	2821	13.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	27.02%	
50) Toluene-d8	10.33	98	6492	7.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	14.08%	
62) 4-Bromofluorobenzene	12.62	95	6157	16.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	33.94%	
Target Compounds						
					Qvalue	
16) Acetone	4.15	43	13825	436.522	ug/l	93
17) Carbon Disulfide	4.39	76	5548	8.509	ug/l #	92
20) Methylene Chloride	4.93	84	2846	6.588	ug/l #	90
25) 2-Butanone	7.18	43	2869	52.891	ug/l #	88
44) Trichloroethene	9.09	130	718	2.626	ug/l	98
64) Tetrachloroethene	10.87	164	171372	698.754	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112818\
Data File : VW007121.D
Acq On : 28 Nov 2018 18:31
Operator : VAY/AP
Sample : J6082-01
Misc : 7.24G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S49-0490

Quant Time: Nov 29 13:26:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

