

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112818\
 Data File : VW007122.D
 Acq On : 28 Nov 2018 18:57
 Operator : VAY/AP
 Sample : J6082-01RE
 Misc : 7.07G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S49-0490RE

Quant Time: Nov 29 05:11:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	22337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	37812	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	38835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	19832	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	13335	60.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.96%	
35) Dibromofluoromethane	7.89	113	4173	19.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.08%	
50) Toluene-d8	10.32	98	10019	10.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	21.26%	
62) 4-Bromofluorobenzene	12.62	95	6926	18.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	37.34%	
Target Compounds						
16) Acetone	4.15	43	11459	360.909	ug/l	95
17) Carbon Disulfide	4.39	76	2573	3.936	ug/l #	94
20) Methylene Chloride	4.92	84	4951	15.882	ug/l	93
25) 2-Butanone	7.19	43	2710	49.835	ug/l	96
44) Trichloroethene	9.10	130	10270	36.731	ug/l	85
64) Tetrachloroethene	10.87	164	241127	965.148	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112818\
Data File : VW007122.D
Acq On : 28 Nov 2018 18:57
Operator : VAY/AP
Sample : J6082-01RE
Misc : 7.07G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S49-0490RE

Quant Time: Nov 29 05:11:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

