

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW112818\
 Data File : VW007126.D
 Acq On : 28 Nov 2018 20:38
 Operator : VAY/AP
 Sample : J5767-23
 Misc : 6.54G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-02-112818-D

Quant Time: Nov 29 05:51:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 03:23:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	14323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	22314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	18003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	6019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	7238	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	7.90	113	6446	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	10.33	98	23317	41.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.80%	
62) 4-Bromofluorobenzene	12.62	95	7419	33.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.76%	
Target Compounds						
16) Acetone	4.14	43	1779	87.381	ug/l	95
20) Methylene Chloride	4.93	84	2306	9.880	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW112818\
Data File : VW007126.D
Acq On : 28 Nov 2018 20:38
Operator : VAY/AP
Sample : J5767-23
Misc : 6.54G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-02-112818-D

Quant Time: Nov 29 05:51:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W112118S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 03:23:02 2018
Response via : Initial Calibration

