

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW113018\
 Data File : VW007148.D
 Acq On : 30 Nov 2018 18:07
 Operator : SY/AP
 Sample : VIBLK82
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK82

Quant Time: Nov 30 23:18:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM113018S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 30 15:51:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	140059	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	137191	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	68786	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	25607	21.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.96%
7) Chloroethane-d5	2.89	69	18530	21.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.72%
10) 1,1-Dichloroethene-d2	4.02	63	49065	15.81	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.24%
20) 2-Butanone-d5	7.08	46	21220	43.98	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.96%
24) Chloroform-d	7.65	84	60745	20.71	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.84%
26) 1,2-Dichloroethane-d4	8.31	65	36598	21.55	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.20%
29) Benzene-d6	8.27	84	128306	19.68	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.72%
33) 1,2-Dichloropropane-d6	9.27	67	37971	20.33	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.32%
37) Toluene-d8	10.33	98	136001	19.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.88%
38) trans-1,3-Dichloropropene-	10.58	79	17118	18.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.00%
39) 2-Hexanone-d5	10.93	63	17774	40.61	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	81.22%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	34653	20.37	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.48%
61) 1,2-Dichlorobenzene-d4	13.86	152	52513	20.34	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW113018\
Data File : VW007148.D
Acq On : 30 Nov 2018 18:07
Operator : SY/AP
Sample : VIBLK82
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK82

Quant Time: Nov 30 23:18:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM113018S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 30 15:51:51 2018
Response via : Initial Calibration

