

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW120219\
 Data File : VW014198.D
 Acq On : 02 Dec 2019 12:49
 Operator : SY/VA
 Sample : K6063-14
 Misc : 6.11G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BP0

Quant Time: Dec 02 16:35:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 02 16:00:46 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	264975	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	233098	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	93418	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	64803	15.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.32%
7) Chloroethane-d5	2.89	69	68910	20.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.44%
10) 1,1-Dichloroethene-d2	4.01	63	114643	13.21	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.84%
20) 2-Butanone-d5	7.07	46	78863	64.21	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.42%
24) Chloroform-d	7.65	84	205077	25.17	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.68%
26) 1,2-Dichloroethane-d4	8.31	65	114353	27.92	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	111.68%
29) Benzene-d6	8.27	84	396611	23.98	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.92%
33) 1,2-Dichloropropane-d6	9.27	67	131117	27.06	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.24%
37) Toluene-d8	10.32	98	335650	21.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.60%
38) trans-1,3-Dichloropropene-	10.58	79	45774	22.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.24%
39) 2-Hexanone-d5	10.92	63	58564	63.75	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	127.50%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	114620	31.33	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	125.32%#
61) 1,2-Dichlorobenzene-d4	13.85	152	98635	26.28	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.12%

Target Compounds					Ovalue
13) Acetone	4.13	43	5513	4.411 ug/L	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW120219\
Data File : VW014198.D
Acq On : 02 Dec 2019 12:49
Operator : SY/VA
Sample : K6063-14
Misc : 6.11G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BP0

Quant Time: Dec 02 16:35:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111119S.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 02 16:00:46 2019
Response via : Initial Calibration

