

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120221\
 Data File : VW021020.D
 Acq On : 02 Dec 2021 03:28
 Operator : SY/VA
 Sample : M4881-08
 Misc : 5.49g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9E4

Quant Time: Dec 02 07:30:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Dec 02 07:10:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	137625	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	117968	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	55036	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	22729	14.238	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.960%
7) Chloroethane-d5	2.885	69	16003	15.543	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.160%
11) 1,1-Dichloroethene-d2	4.019	63	30966	10.026	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	40.120%#
21) 2-Butanone-d5	7.074	46	16953	42.547	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.100%
24) Chloroform-d	7.653	84	63626	18.348	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.400%
26) 1,2-Dichloroethane-d4	8.305	65	34534	19.620	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.480%
32) Benzene-d6	8.275	84	112260	18.055	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	72.240%
36) 1,2-Dichloropropane-d6	9.274	67	35069	19.807	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.240%
41) Toluene-d8	10.323	98	106118	17.196	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.800%
43) trans-1,3-Dichloroprop...	10.579	79	13996	16.725	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.880%
47) 2-Hexanone-d5	10.921	63	13227	45.045	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.080%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	29049	21.418	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.680%
66) 1,2-Dichlorobenzene-d4	13.847	152	38643	19.700	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.800%
Target Compounds						
13) Acetone	4.129	43	3028	8.512	ug/L	Qvalue 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120221\
Data File : VW021020.D
Acq On : 02 Dec 2021 03:28
Operator : SY/VA
Sample : M4881-08
Misc : 5.49g/10.0mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9E4

Quant Time: Dec 02 07:30:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM111521SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Dec 02 07:10:34 2021
Response via : Initial Calibration

