

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021093.D
 Acq On : 04 Dec 2021 03:40
 Operator : SY/VA
 Sample : M4883-05
 Misc : 5.64g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9G1

Quant Time: Dec 04 05:45:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 04:50:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	107399	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	102165	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	50925	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.348	65	50994	34.079	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 136.320%	
7) Chloroethane-d5	2.885	69	30884	30.832	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 123.320%	
11) 1,1-Dichloroethene-d2	4.019	63	52348	21.899	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 87.600%	
21) 2-Butanone-d5	7.098	46	13179	38.613	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	= 77.220%	
24) Chloroform-d	7.652	84	77827	28.936	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 115.760%	
26) 1,2-Dichloroethane-d4	8.305	65	40318	27.400	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 109.600%	
32) Benzene-d6	8.274	84	153911	29.464	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 117.840%	
36) 1,2-Dichloropropane-d6	9.274	67	42935	29.576	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 118.320%	
41) Toluene-d8	10.323	98	149920	27.989	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 111.960%	
43) trans-1,3-Dichloroprop...	10.579	79	17348	26.318	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 105.280%	
47) 2-Hexanone-d5	10.926	63	10848	39.058	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 78.120%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	27618	20.564	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 82.240%	
66) 1,2-Dichlorobenzene-d4	13.852	152	45693	25.042	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 100.160%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021093.D
Acq On : 04 Dec 2021 03:40
Operator : SY/VA
Sample : M4883-05
Misc : 5.64g/10.0mL/MSVOA_W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9G1

Quant Time: Dec 04 05:45:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 04:50:24 2021
Response via : Initial Calibration

