

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021101.D
 Acq On : 04 Dec 2021 06:51
 Operator : SY/VA
 Sample : M4883-13
 Misc : 2.87g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9G9

Quant Time: Dec 04 07:10:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 04:50:24 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	30679	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.627	117	19515	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	4854	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	19915	46.591	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 186.360%#	
7) Chloroethane-d5	2.885	69	15836	55.344	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 221.360%#	
11) 1,1-Dichloroethene-d2	4.019	63	23950	35.074	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 140.280%#	
21) 2-Butanone-d5	7.086	46	10487	107.562	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 215.120%#	
24) Chloroform-d	7.653	84	41733	54.318	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 217.280%#	
26) 1,2-Dichloroethane-d4	8.305	65	24166	57.494	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 229.960%#	
32) Benzene-d6	8.281	84	56501	56.625	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 226.480%#	
36) 1,2-Dichloropropane-d6	9.274	67	23281	83.958	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 335.840%#	
41) Toluene-d8	10.323	98	40888	39.963	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 159.840%#	
43) trans-1,3-Dichloroprop...	10.579	79	5832	46.319	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 185.280%#	
47) 2-Hexanone-d5	10.926	63	6636	125.082	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 250.160%#	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	15974	62.269	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 249.080%#	
66) 1,2-Dichlorobenzene-d4	13.853	152	5214	29.980	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 119.920%	
Target Compounds						Qvalue
13) Acetone	4.141	43	1368	15.404	ug/L	75
16) Methylene chloride	4.915	84	3181	6.581	ug/L	94
27) 1,2-Dichloroethane	8.402	62	1404	2.844	ug/L	# 93
34) Trichloroethene	9.091	95	43140	140.906	ug/L	91
42) Toluene	10.390	91	1257	0.992	ug/L	95
46) Tetrachloroethene	10.865	164	8875	31.657	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021101.D
Acq On : 04 Dec 2021 06:51
Operator : SY/VA
Sample : M4883-13
Misc : 2.87g/10.0mL/MSVOA_W/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9G9

Quant Time: Dec 04 07:10:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 04:50:24 2021
Response via : Initial Calibration

