

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021107.D
 Acq On : 04 Dec 2021 09:15
 Operator : SY/VA
 Sample : M4885-02
 Misc : 5.01g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9K9

Quant Time: Dec 06 00:48:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 04:50:24 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	100398	25.000	ug/L	# 0.00
28) Chlorobenzene-d5		11.634	117	97834	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.560	152	49427	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.355	65	40299	28.809	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	115.240%	
7) Chloroethane-d5		2.885	69	27596	29.471	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	117.880%	
11) 1,1-Dichloroethene-d2		4.019	63	46621	20.863	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.440%	
21) 2-Butanone-d5		7.086	46	17683	55.422	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.840%	
24) Chloroform-d		7.653	84	75010	29.833	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	119.320%	
26) 1,2-Dichloroethane-d4		8.311	65	43071	31.312	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	125.240%	
32) Benzene-d6		8.281	84	147254	29.437	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	117.760%	
36) 1,2-Dichloropropane-d6		9.274	67	42846	30.821	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	123.280%#	
41) Toluene-d8		10.323	98	138914	27.083	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	108.320%	
43) trans-1,3-Dichloroprop...		10.579	79	17477	27.688	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	110.760%	
47) 2-Hexanone-d5		10.927	63	13531	50.874	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.740%	
56) 1,1,2,2-Tetrachloroeth...		12.695	84	33097	25.735	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.920%	
66) 1,2-Dichlorobenzene-d4		13.853	152	48550	27.414	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.640%	
Target Compounds							
34) Trichloroethene		9.092	95	3141	2.046	ug/L	84
46) Tetrachloroethene		10.866	164	1122	0.798	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021107.D
Acq On : 04 Dec 2021 09:15
Operator : SY/VA
Sample : M4885-02
Misc : 5.01g/10.0mL/MSVOA_W/SOIL
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9K9

Quant Time: Dec 06 00:48:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 04:50:24 2021
Response via : Initial Calibration

