

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
 Data File : VW021109.D
 Acq On : 04 Dec 2021 10:03
 Operator : SY/VA
 Sample : M4885-04
 Misc : 5.72g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW9L1

Quant Time: Dec 06 00:48:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 04 04:50:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	97624	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	89785	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	40567	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	45488	33.443	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 133.760%		
7) Chloroethane-d5	2.885	69	30076	33.032	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 132.120%		
11) 1,1-Dichloroethene-d2	4.013	63	48655	22.392	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 89.560%		
21) 2-Butanone-d5	7.080	46	19625	63.256	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 126.520%		
24) Chloroform-d	7.647	84	77268	31.604	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 126.400%		
26) 1,2-Dichloroethane-d4	8.311	65	45978	34.376	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 137.520%#		
32) Benzene-d6	8.275	84	150172	32.712	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 130.840%		
36) 1,2-Dichloropropane-d6	9.275	67	43636	34.204	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 136.800%#		
41) Toluene-d8	10.323	98	137747	29.263	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 117.040%		
43) trans-1,3-Dichloroprop...	10.579	79	18149	31.330	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 125.320%		
47) 2-Hexanone-d5	10.921	63	15351	62.891	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 125.780%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	33892	28.716	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 114.880%		
66) 1,2-Dichlorobenzene-d4	13.853	152	37662	25.911	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 103.640%		
Target Compounds						Qvalue
13) Acetone	4.141	43	1092	3.864	ug/L	93
46) Tetrachloroethene	10.860	164	3192	2.475	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW120321\
Data File : VW021109.D
Acq On : 04 Dec 2021 10:03
Operator : SY/VA
Sample : M4885-04
Misc : 5.72g/10.0mL/MSVOA_W/SOIL
ALS Vial : 59 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW9L1

Quant Time: Dec 06 00:48:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120321SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 04 04:50:24 2021
Response via : Initial Calibration

